

Analysenzertifikat

Certificate of Analysis

SERION ELISA *classic*

Zusatzreagenz / Supplementary Reagent

SERION RF Absorbent

Best.-Nr. / Order No.	Z200
Chargen-Nr. / Lot No.	CR0015
IFU-Version / IFU version	12-25/06-2
Herstelldatum / Manufacturing date	2026-01
Verfallsdatum / Expiry Date	2028-01-31
Anwendungsgebiet / Field of Application	Entfernung von Rheumafaktoren durch Absorption von 15 mg/ml IgG Removal of rheumatoid factor by absorption of 15 mg/ml IgG
Material / Material	Ziegenserum / Goat serum
Konservierungsmittel / Preservation	0,09% NaN ₃
Lagerung / Storage	2°C - 8°C
Gebrauchsverdünnung / Working Dilution	gebrauchsfertig / ready to use



Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: info@virion-serion.de
 Internet: www.virion-serion.de